

Standard Operating Procedure (SOP)

Title: Double Polyclonal Enzyme-Linked Immunosorbent Assay (ELISA) for the detection of *Renibacterium salmoninarum*

Number: BACT-2

Version: 02

Approval:	Date:
Bacteriology Supervisor:	March 28, 2012

Effective Date:	March 28, 2012
Document Replaced:	V1
Reason for Revision:	Mistake in washing procedure
Area of Application:	Bacteriology

Purpose: To detect *Renibacterium salmoninarum* soluble antigen in fish tissues or fluids

Sections:

- I. Overview
- II. References
- III. Important considerations
- IV. Procedures to do before starting ELISA
- V. ELISA Procedure
 - Appendix A—Reagent preparation
 - Appendix B—Checkerboard titration
 - Appendix C—Preparing samples
 - Appendix D—Description of controls

I. Overview

The *Renibacterium salmoninarum* enzyme-linked immunosorbent assay (ELISA) detects the presence and amount of *R. salmoninarum* antigens in fish tissues or ovarian fluid. The basic steps are:

1. Coat ELISA plate with coating antibody (anti-*R. salmoninarum* antibody) overnight
2. Wash and add processed tissue or ovarian fluid sample
3. Wash and add the conjugate (horseradish peroxidase (HRP)-anti-*R. salmoninarum*)
4. Wash and add substrate-chromogen
5. Add stop solution and read absorbance (optical density or OD)

II. References

- Meyers T.R., S. Short, C. Farrington, K. Lipson, H.J. Geiger, and R. Gates. 1993. Establishment of a negative-positive threshold optical density value for the enzyme-linked immunosorbent assay (ELISA) to detect soluble antigen of *Renibacterium salmoninarum* in Alaskan Pacific salmon. *Diseases of Aquatic Organisms* 16:191-197
- Pascho, R. J., and D. Mulcahy. 1987. Enzyme-linked immunosorbent assay for a soluble fraction of *Renibacterium salmoninarum*, the causative agent of salmonid bacterial kidney disease. *Canadian Journal of Fisheries and Aquatic Sciences* 44: 183-191.
- Pascho, R.J., D.G. Elliott, R.W. Mallett, and D. Mulcahy. 1987. Comparison of five techniques for the detection of *Renibacterium salmoninarum* in adult coho salmon. *Transactions of the American Fisheries Society* 116:882-890.
- Pascho, R. J., D. G. Elliot, and J. M. Streufert. 1991. Broodstock segregation of spring Chinook salmon *Oncorhynchus tshawytscha* by use of the enzyme-linked immunosorbent assay (ELISA) and the fluorescent antibody techniques (FAT) affects the prevalence and levels of *Renibacterium salmoninarum* infection in progeny. *Diseases of Aquatic Organisms* 12:25-40.

III. Important considerations

Glassware: It is advisable to dedicate a separate set of glassware for use in preparing reagents for ELISA. Proteins can remain on container surfaces after conventional detergent washing and autoclaving. Disposable containers can also be used to decrease risk of contamination.

Checkerboard: See Appendix B for complete checkerboard instructions. The working concentrations of the coating antibody (anti-*R. salmoninarum*) and the conjugate (HRP-anti-*R. salmoninarum*) must first be determined by checkerboard titration. Typical concentrations are around 1 µg per mL for the coating antibody and 1:4,000 (v/v) for the conjugate antibody.

Samples: See Appendix C for complete processing instructions. Homogenized fish tissues and ovarian fluid should be diluted with PBS-T20 and stored in 2 mL screw cap centrifuge tubes with O-rings at -70°C until needed. To prepare samples for the ELISA, heat each sample at 100°C for 15 min, and then centrifuge 8,000 to 10,000 x g for 10 minutes at 4°C. If the samples were

prepared earlier and frozen, thaw before heating. Heated and centrifuged samples can be stored overnight at 4°C, or frozen at -70°C for later testing.

Plates: The WFRC uses the 96 well EIA/RIA clear flat bottom plate from Corning Inc. (#3590).

IV. Procedures to do before starting ELISA

1. Prepare controls (See Appendix A).

Positive controls:

- 1 - 1:100
- 2 - 1:1000
- 3 - 1:2000
- 4 - 1:5000

Negative control

2. Assemble reagents (See Appendix A).

- ELISA plate(s)
- Wash solution
- PBS-T20
- Coating buffer
- Coating antibody
- Milk diluent
- HRP-conjugate antibody
- Substrate- chromogen
- Stop solution

V. ELISA Procedure

Day 1

- 1. Prepare microplate map.** A control plate with replicates of blank (BL), conjugate control (CC), substrate-chromogen control (SC), negative control (NC), and all concentrations of positive control wells should be done for every four plates processed or every new run of ELISA. (See example control plate below). Sequentially number each plate.

Example control plate

	1	2	3	4	5	6	7	8	9	10	11	12
A	NC	CC	PC 1:100	PC 1:2000		Sample	Sample	Sample	Sample	Sample	Sample	Sample
B	BL	CC	PC 1:100	PC 1:2000		Sample	Sample	Sample	Sample	Sample	Sample	Sample
C	NC	CC	PC 1:100	PC 1:2000		Sample	Sample	Sample	Sample	Sample	Sample	Sample
D	BL	CC	PC 1:100	PC 1:2000		Sample	Sample	Sample	Sample	Sample	Sample	Sample
E	NC	SC	PC 1:1000	PC 1:5000		Sample	Sample	Sample	Sample	Sample	Sample	Sample
F	BL	SC	PC 1:1000	PC 1:5000		Sample	Sample	Sample	Sample	Sample	Sample	Sample
G	NC	SC	PC 1:1000	PC 1:5000		Sample	Sample	Sample	Sample	Sample	Sample	Sample
H	BL	SC	PC 1:1000	PC 1:5000		Sample	Sample	Sample	Sample	Sample	Sample	Sample

NC= negative control, BL=blank well, CC=coat control, SC=substrate-chromagen control, PC=positive control

- 2. Calculate the amount of coating buffer and coating antibody needed to coat each microplate.**

Prepare an excess volume of coating antibody and buffer to compensate for solution loss.

Example 1

Calculate total volume of coating antibody at 1 $\mu\text{g}/\text{mL}$ [§] necessary to apply 200 μL of antibody solution to the wells of 5 microplates:

$$(5 \text{ plates}) (96 \text{ wells/plate}) (200 \mu\text{L}/\text{well}) = 96,000 \mu\text{L} = 96 \text{ mL}$$

To compensate for loss, prepare 105 mL of diluted coating antibody at 1 $\mu\text{g}/\text{mL}$

$$V_i * (1000 \mu\text{g}/\text{mL}) = 105 \text{ mL} * (1 \mu\text{g}/\text{mL})$$

$$V_i = 0.105 \text{ mL} = 105 \mu\text{L}$$

Combine 105 μL concentrated coating antibody with 105mL coating buffer.

[§]Concentration determined by checkerboard

3. **Dilute the coating antibody in coating buffer.**
4. **Place 200 μ L of coating buffer only in the conjugate control (CC) and substrate- chromogen control (SC) wells.**
5. **Coat all other wells in the microplate with 200 μ l diluted coating antibody.**
6. **Seal each plate with an adhesive plate sealer.**
7. **Place each plate in a humid chamber to incubate at 4°C for 16+ hrs.**
8. **Prepare wash buffer solution for Day 2.** Wash buffer can be stored at 4°C overnight, but should remain at room temperature during the ELISA.
9. **Prepare test samples for Day 2.** Heat each sample at 100°C for 15 min, centrifuge 8,000 to 10,000 $\times g$ for 10 minutes at 4°C. If the samples were prepared earlier and frozen, thaw before heating. Heated and centrifuged samples can be stored overnight at 4°C (see Appendix C). Remember to also prepare positive and negative controls along with test samples (see Appendix D).

Day 2

1. **Wash wells five times with wash buffer.** Carefully remove the plate sealer, fill each well with wash buffer, and allow the buffer to soak for 30 seconds before removing the solution. Shake out any excess buffer left over after all five washes are complete. An autowasher is highly recommended for this step.
2. **Add 200 μ L of only PBS-T20 into blank (BL), conjugate control (CC), and substrate-chromogen control (SC) wells.**
3. **Add 200 μ L of prepared negative tissue or ovarian fluid sample into the negative control (NC) wells.** Seal negative control wells with a strip of adhesive plate sealer before loading positive controls.
4. **Add 200 μ L of each positive control into the appropriate wells.** Seal positive control wells with another strip of adhesive plate sealer before loading test samples.
5. **Add 200 μ L aliquots of each test sample (supernatant only) in the appropriate wells.** Record any discrepancies on your microplate map (e.g. sample wells that did not receive a full 200 μ L or samples with tissue debris in the supernatant). Samples are usually tested in duplicate if volume is sufficient.
6. **Cover each plate with a full adhesive plate sealer.** Write time of completion of the loading of each plate on the sealer flap of that plate and on the microplate map.

7. **Incubate for 3 h at 25°C in a humid chamber.**
8. **Wash wells five times as described in step 1.**
9. **Calculate amount of HRP-conjugate antibody and milk diluent needed for 200µL/well.** The HRP-conjugate should be diluted to the concentration previously obtained in the Checkerboard titration (See Appendix B).

Example 2

Calculate the volume of a 1:4,000 dilution^s of the HRP-conjugate antibody for 5 microplates.

$$(5 \text{ plates}) (96 \text{ wells/plate}) (200 \text{ } \mu\text{L/well}) = 96,000 \text{ } \mu\text{L} = 96 \text{ mL}$$

Make 105 mL to account for loss

$$\begin{aligned} 1/4000 &= x/105 \text{ mL} \\ x &= 0.0263 \text{ mL} = 26.3 \text{ } \mu\text{L} \end{aligned}$$

Combine 26.3 µL of the stock HRP-conjugate with 105 mL milk diluent
§Dilution determined by checkerboard

10. **Add the appropriate amount of HRP- conjugated antibody to 1X blocking solution (also called milk diluent).**
11. **Add 200 µL of the diluted conjugate to the appropriate wells.** Substrate-chromogen control wells (SC) receive 200 µL of milk diluent **without** HRP-conjugate.
12. **Seal each plate with an adhesive plate sealer.**
13. **Incubate in a humid chamber for 2 h at 25°C.**
14. **Wash wells five times as described in step 1.**
15. **Calculate amount of substrate-chromogen needed for 200µL/well.** The substrate-chromogen solutions should be at room temperature before use. Remove from refrigerator during HRP-conjugate incubation step.

Example 3

*Calculate the volumes needed of substrate-chromogen for 5 microplates:
equal volumes ABTS solutions A and B necessary for 5 microplates.*

$$(5 \text{ plates}) * (96 \text{ wells/plate}) * (200 \mu\text{L/well}) = 96,000 \mu\text{L} = 96 \text{ mL}$$

Make 105 mL to compensate for loss

$$52.5 \text{ mL of A} + 52.5 \text{ mL of B} = 105 \text{ mL}$$

16. **For each plate, set a timer for 20 min (but do not start).** The timing of the substrate-chromogen reaction is critical. The reaction must be stopped after exactly 20 minutes.
17. **Prepare substrate-chromogen solution and add 200 μL to each well.** Start the appropriate timer immediately when you begin to apply the solution. Fill all plates in numerical order.
18. **Immediately after all the wells have received the substrate-chromogen solution, seal the plate and put it in the humid chamber at 37°C.**
19. **Calculate the amount of stop solution needed.** Dilute 5x concentrate 4:1 and use 50 μL of 4x solution per well.

Example 4

Calculate the volume of 4X stop solution needed for 5 microplates.

$$(5 \text{ plates}) (96 \text{ wells/plate}) (50 \mu\text{L/well}) = 24,000 \mu\text{L} = 24 \text{ mL}$$

Make 34 mL to compensate for loss

$$5X * V_i = 4X * 34$$

$$V_i = 27.2 \text{ mL concentrate}$$

Combine 6.8 mL of nH₂O with 27.2 mL of concentrated stop solution to get 34 mL total

20. **Apply 50 μL of the stop solution to each well immediately after 20 minute incubation period is complete.** Wells should receive the stop solution in the exact sequence used to apply the substrate-chromogen solution.
21. **Immediately measure absorbance at 405nm wavelength in a microplate reader.** Wipe any condensation off the bottom of the plate with a kimwipe before reading.

Data Interpretation

Note: The optical density values for control and test reagents may vary between laboratories using different equipment or manufacturer models. Here we will provide general guidelines for interpreting data.

1. Quality Control

- a. Variation between replicates (coefficient of variation) should be less than 10%.
- b. The blank wells should have low OD values (e.g. < 0.08 OD).
- c. The SC and CC controls should have similar values and low OD values (e.g. <0.08 OD).

2. Negative-Positive Threshold Determination

Note: In some circumstances, such as when fish are to be transported to a watershed or hatchery that has no known *R. salmoninarum* in resident fish stocks, a conservative negative-positive threshold value is often used. However, the use of a conservative threshold increases the likelihood of false-positive ELISA OD values, and may not be desirable for monitoring of fish populations that are known to be *R. salmoninarum*-positive. In circumstances such as the culling of gametes from fish in *R. salmoninarum*-positive populations to reduce vertical transmission of the bacterium, a less conservative threshold OD is often used, although this increases the likelihood of false-negative results. Please consult Meyers et al. (1993) for further discussion of ELISA threshold values.

- a. A conservative negative-positive threshold is often calculated as the mean of the negative control wells (NC) plus two standard deviations. These metrics are easily calculated within Microsoft Excel.

Example 5

NC 1 = 0.072 OD

NC 2 = 0.069 OD

NC 3 = 0.067 OD

Then the mean is 0.069 and the standard deviation is 0.003
The positive – negative threshold is $0.069 + (2 \times 0.003) = 0.074$

Appendix A: Reagent Preparation

Notes: See section II on glassware; all protocols use nanopure or reagent grade water

PBS-T20: Phosphate-buffered saline, pH 7.4, 0.05% (v/v) Tween-20 & 0.01% (w/v) Thimerosal

For 1 liter:

NaCl	8.00 g
KH ₂ PO ₄	0.20 g
Na ₂ HPO ₄	1.09 g
KCl	0.20 g
Thimerosal	0.10 g
<i>-Confirm pH is 7.4</i>	
Tween-20	0.5 mL

PC: *Renibacterium salmoninarum* Postive Control BacTrace®

- Purchase commercially (KPL[§] part # 50-96-91)
 - Reconstitute 5 vials of positive control with 5 mL of PBS-T20*
 - Prepare PC dilutions
 - 1:100 Add 5 mL of reconstituted antigen to 495 mL of PBS-T20
 - 1:1000 Add 40 mL of 1:100 to 360 mL PBS-T20
 - 1:2000 Add 20 mL of 1:100 to 380 mL PBS-T20
 - 1:5000 Add 8 mL of 1:100 to 392 mL of PBS-T20
- Note: Dilutions should give a standard curve between OD₄₀₅ 0.10 and 2.25*
- Aliquot into 2 mL screw-cap micro tubes* and freeze at -70°C
- *Notes: Rehydration instructions differ from manufacturer
Different color caps can be used to distinguish positive controls
2 mL screw cap tubes with O-rings that can withstand boiling*

NC: Negative Control

- Confirmed negative kidney or ovarian fluid for negative control

Coating Buffer: Coating Solution Concentrate (10X)

- Purchase commercially from KPL[§] (Part # 50-84-01)
- Store at 4°C after opening and is usually stable for a minimum of one year
- Dilute to 1X concentration for use (1 mL concentrate to 9 mL water)

Coating Antibody: Affinity Purified Antibody to *Renibacterium salmoninarum* BacTrace®
Antibodies produced in Goat

- Purchase commercially from KPL[§] (part #01-96-91)
- Vial contains 1.0 mg of lyophilized immunoglobulin
- Rehydrate antibody in 1 mL of 50% sterile-filtered (0.2 µm) glycerol
- Store at 4°C, usually stable for a minimum of one year

Appendix A. continued**Wash Solution:** Wash Solution Concentrate (20X)

- Purchase commercially from KPL[§] (part # 50-63-01)
- Store concentrate at 4°C after opening; usually stable for a minimum of one year
- Dilute according to manufacturer's instructions (1 mL concentrate to 19 mL water)
- Prepare fresh for each ELISA

Blocking Solution: 20X Blocking Solution Concentrate / Milk Diluent

- Purchase commercially from KPL[§] (Part # 50-82-00)
- Store concentrate at 4°C after opening; usually stable for a minimum of one year
- Dilute following manufacturer's instructions (1 mL Diluent/Block + 19 mL water)

Antibody Conjugate: Peroxidase-Labeled Affinity Purified Antibody to *Renibacterium salmoninarum* BacTrace® Antibodies produced in Goat

- Purchase from KPL[§] (Part # 04-96-91)
- Vial contains 0.1 mg
- Rehydrate according to manufacturer's instructions (Procedure A: 1 ml of 50% glycerol)
- Store at 4°C in a dark or foil wrapped tube; usually stable for a minimum of one year

Substrate-Chromogen: ABTS® Peroxidase Substrate System (2 Component)

- Purchase ABTS Peroxidase Solution A from KPL[§] (Part # 50-64-02)
2,2'-azino-di[3-ethyl-benzthiazoline sulfonate in a 0.3g/L glycine/citric acid buffer
- Purchase ABTS Peroxidase Solution B from KPL[§] (Part # 50-65-02)
0.02% hydrogen peroxide in a proprietary buffer
- Store in dark bottles at 4°C

Stop Solution: ABTS® Peroxidase Stop Solution

- Purchase from KPL[§] (Part # 50-85-02)
- Store at room temperature; stable for a minimum of one year.
- Combine 4 parts stop solution with 1 part water immediately before use*
**Note dilution differs from manufacturer's recommendation*

§Kirkegaard and Perry Laboratories Inc. (KPL)
910 Clopper Road, Gaithersburg, Maryland 20878 USA
Ph: 800.638.3167; www.kpl.com

Appendix B: Checkerboard Titration

Purpose: to determine the working concentrations of coating antibody and conjugate antibody

Background

Each time new vials of the coating antibody or conjugate are rehydrated, slight variations between lots of the manufacturer's antibody preparations, the rehydration procedure, etc. may affect the amounts of those stock preparations used in the ELISA. To determine the optimal (working) concentration of each reagent, several dilutions of antibody must be evaluated together with a prescribed set of test samples. From these data one can choose a coating antibody-conjugate combination that produces acceptable absorbance values. If rehydrated antibodies have not been used for an ELISA for several months, it is wise to do another checkerboard titration before testing new tissue or ovarian fluid samples.

Procedure

Day 1

1. **Prepare checkerboard plate map.** Suggested format for evaluating one concentration of the coating antibody with four concentrations of conjugate antibody is shown below. A separate microplate is used for each concentration of coating antibody tested.
 - a. Suggested concentrations for **conjugate antibody** are 1:2,500, 1:3,000, 1:4,000, and 1:4,500 (v/v) dilutions.
 - b. The optimal **coating antibody** is usually 1.0 µg/mL but 1.5 and 2.0 µg/mL concentrations can be tested on separate plates.
 - c. **Controls** include: blank wells (BL), the negative control (NC), the conjugate control (CC), substrate-chromogen control (SC), and the 4 positive control (PC) dilutions (1:1000 - 1:5,000). The controls are described in Appendix D.

Checkerboard plate: Each quadrant uses a different concentration of conjugate antibody.

	1	2	3	4	5	6	7	8	9	10	11	12
A	NC	BL	SC	CC	PC 1:100	PC 1:2000	NC	BL	SC	CC	PC 1:100	PC 1:2000
B	NC	BL	SC	CC	PC 1:100	PC 1:2000	NC	BL	SC	CC	PC 1:100	PC 1:2000
C	NC	BL	SC	CC	PC 1:1000	PC 1:5000	NC	BL	SC	CC	PC 1:1000	PC 1:5000
D	NC	BL	SC	CC	PC 1:1000	PC 1:5000	NC	BL	SC	CC	PC 1:1000	PC 1:5000
E	NC	BL	SC	CC	PC 1:100	PC 1:2000	NC	BL	SC	CC	PC 1:100	PC 1:2000
F	NC	BL	SC	CC	PC 1:100	PC 1:2000	NC	BL	SC	CC	PC 1:100	PC 1:2000
G	NC	BL	SC	CC	PC 1:1000	PC 1:5000	NC	BL	SC	CC	PC 1:1000	PC 1:5000
H	NC	BL	SC	CC	PC 1:1000	PC 1:5000	NC	BL	SC	CC	PC 1:1000	PC 1:5000

Appendix B continued.

2. **Rehydrate lyophilized antibody and conjugate according to manufacturer's instructions.** See Appendix A for more information.
3. **Calculate the amount of coating buffer and coating antibody needed for the checkerboard plate.** Prepare an excess volume of coating antibody to compensate for solution lost during preparation and pipetting.

Example 1

Calculation of the total volume of coating antibody at 1 µg/mL is necessary to apply 200 µL of antibody to the wells of 1 microplate.

$$(64 \text{ wells/plate}) (200 \text{ µL/well}) = 12,800 \text{ µL} = 12.8 \text{ mL}$$

Prepare 16 mL coating antibody at 1 µg/mL

$$X (1,000 \text{ µg/mL}) = 16,000 \text{ µL} (1 \text{ µg/mL})$$

$$X = 16 \text{ µL}$$

Combine 16 µL concentrated anti-R. salmoninarum coating antibody with 16 mL coating buffer.

4. **Dilute the coating antibody in coating buffer.**
5. **Place 200µL of coating buffer only in the conjugate control (CC) and substrate-chromogen control (SC) wells.**
6. **Coat all other wells in the microplate with diluted anti-R. salmoninarum coating antibody and coating buffer.**
7. **Seal each plate with an adhesive plate sealer.**
8. **Place each plate in a humid chamber to incubate at 4°C for 16+ h.**
9. **Prepare wash buffer solution for day 2.** Wash buffer can be stored at 4°C overnight, but should remain at room temperature during the ELISA.
10. **Prepare controls for Day 2.** Heat controls at 100°C for 15 min, centrifuge 8,000 to 10,000 x g for 10 minutes at 4°C. If the samples were prepared earlier and frozen, thaw before heating. Processed and heated controls can be stored overnight at 4°C (see Appendix C).

Appendix B continued.**Day 2**

1. **Wash wells five times with wash buffer.** Fill each well with wash buffer, allow the buffer to soak for 30 second before removing the solution. Shake out any excess buffer left over after all five washes are complete. An autowasher is highly recommended for this step.
2. **Add 200 μ L of only PBS-T20 into blank (BL), conjugate control (CC), and substrate-chromogen control (SC) wells.**
3. **Add 200 μ L of prepared negative tissue or ovarian fluid sample into the negative control (NC) wells.** Seal control wells with a strip of adhesive plate sealer before loading positive controls.
4. **Add 200 μ L of each positive control into the appropriate wells.** Seal positive control wells with another strip of adhesive plate sealer before loading test samples.
 - Positive controls:
 - 1 - 1:100
 - 2 - 1:1000
 - 3 - 1:2000
 - 4 - 1:5000
5. **Cover plate with a full adhesive plate sealer.**
6. **Write time of completion on the sealer flap on each plate and on the template sheet.**
7. **Incubate for 3 h at 25°C in a humid chamber.**
8. **Wash wells five times as described in step 1.**
9. **Calculate amount of conjugate antibody and blocking solution needed for 200 μ L/well, for four concentrations (1:2,500, 1:3,000, 1:4,000, and 1:4,500 (v/v)).**

Example 2

Calculate the volume of a 1:2,500, dilution of the HRP-conjugate antibody for 1 microplate.

$$(20 \text{ wells}) (200 \mu\text{L/well}) = 4,000 \mu\text{L} = 4 \text{ mL}$$

Make ~ 6 mL extra, or 10 mL.

$$1/2500 = X/10 \text{ mL}$$

$$X = 0.0040 \text{ mL} = 4.0 \mu\text{L}$$

Combine 4.0 μ L of the stock conjugate antibody with 10 mL blocking solution

Appendix B continued.

- 10. Prepare the blocking solution and add the appropriate amount of conjugate antibody, for each of the four concentrations.**
- 11. Add 200 μ L of the diluted conjugate for each concentration to the appropriate wells in each quadrant.**
- 12. Substrate-chromogen control wells (SC) receive 200 μ L of blocking solution without conjugated antibody.**
- 13. Seal plate with an adhesive plate sealer and incubate in a humid chamber for 2 h at 25°C.**
- 14. Wash wells five times as described in step 1.**
- 15. Calculate amount of substrate-chromogen needed for 200 μ L/well.**

Example 3

Calculate the volume substrate-chromogen needed (equal volumes ABTS solutions A&B) necessary for 1 microplate.

$$(1 \text{ plates}) (96 \text{ wells/plate}) (200 \mu\text{L/well}) = 19,200 \mu\text{L} = 19.2 \text{ mL}$$

Again, make ~10 mL extra, or 30 mL.

$$\frac{15 \text{ mL of A} + 15 \text{ mL of B}}{= 30 \text{ mL}}$$

Combine 15 mL of solution A and 15 mL of solution B to get 30 mL total.

- 16. Set a timer for 20 min (but don't start).**
- 17. Prepare substrate-chromogen solution and add 200 μ L to each well.** Start the appropriate timer immediately when you begin to apply the solution. Fill all plates in numerical order. The timing of the substrate-chromogen reaction is critical. The reaction must be stopped after exactly 20 minutes.
- 18. Seal the plate immediately after all the wells have received the substrate-chromogen solution.**
- 19. Put plates in the humid chamber at 37°C.**
- 20. Prepare 4X stop solution.**

Appendix B continued.**Example 4**

Calculate the volume of 4X stop solution needed for 1 microplates.

$$(1 \text{ plates}) (96 \text{ wells/plate}) (50 \mu\text{L/well}) = 4,800 \mu\text{L} = 4.8 \text{ mL}$$

Make 15 mL to compensate for loss

$$5X * V_i = 4X * 15$$

$$V_i = 12 \text{ mL concentrate}$$

Combine 3 mL of nH₂O with 12 mL of concentrated stop solution to get 15 mL total

21. **Begin to apply 50 μL of the 4X stop solution to each well immediately after the incubation period is complete.** Wells should receive the stop solution in the exact sequence used to apply the substrate-chromogen solution.
22. **Place plate in microplate reader and measure absorbances at 405 nm.** Wipe any condensation off the bottom of the plate with a kimwipe before putting into reader.

VIII. Appendix C: Collection and Processing of Tissue or Body Fluid Samples for ELISA

A. Sample Collection and Handling

- Field containers should be labeled, sterile, stomacher-type bags with a top closure, for minimal handling.
- All samples should be treated similarly, e.g., dissect fish, store on ice, freeze-thaw, add diluent and homogenize/transfer sample into microcentrifuge tube, etc.
- Store processed tissues at -70°C .
- Use PBS-T20 as sample buffer.
- Recommended sample tubes are 2 mL screw-cap microcentrifuge tubes with O-rings that can withstand boiling.
- All fish taken from a group of fish must be processed identically. You cannot change the dilution of test samples within, or among, test groups.

B. Sample collection from adult fish

- Kidney: Collect a 2 to 5 gram sample; when sampling the kidney, it is recommended that this sample consist of a pool of tissue pieces from the anterior, mid, and posterior kidney.
- Ovarian fluid: collect approximately 1mL.
- Keep tissue and body fluid samples on ice during collection.
- Take care to avoid cross-contamination between fish and contamination of tissue sample from body fluids

C. Sample collection from juvenile fish

- Kidney: Remove the entire kidney and spleen from each fish. It is preferable to test tissues from individual fish. Two-fish pools are recommended when an insufficient amount of tissue is recovered from an individual fish. If you cannot adhere to these guidelines, then the fish are too small for testing.
- Juvenile fish are often collected as whole fish and dissected upon return to the laboratory.
- As a minimum, need approximately 0.08 g of sample for testing.
- Keep fish or tissues on ice during collection and processing.

D. Sample Processing

– Important Considerations

- See section II (Important Considerations: Glassware)
- Avoid rupturing the stomach or intestine. Commercial fish food frequently contains antigens of *R. salmoninarum* which can be detected by the ELISA, resulting in a false positive reading.
- When tissues are removed in the laboratory, instruments should be flamed between fish in a given group, and mechanically cleaned with detergent between groups of fish. For field sampling, use disposable instruments when possible to remove organs, or have a sufficient number of pre-cleaned instruments so that there is a separate instrument for each fish. Instruments can be cleaned using detergent, dried and autoclaved.
- Store processed tissues at -70°C before testing.

Appendix C continued.**– Processing kidney or spleen**

- Adult fish: 1:4 (w/v); 1 part tissue + 3 parts PBS-T20.
 - Homogenize with print rollers or stomacher in bag.
- Juvenile fish: Recommend 1:4 (w/v) dilution or use 1:8 (w/v); 1 part tissue + 7 parts PBS-T20.
 - Place tissue from individual fish directly into a pre-weighed microcentrifuge tube and homogenize with a stirring rod (Sarstedt, Catalogue Number 81.970).
- Dilute and homogenize tissue.
 - This material can then be stored at -70°C at this point
- Sample are heated at 100°C for 15 min. and centrifuged (8,000-10,000 x g for 10 minutes) to remove debris. The supernatant is used in the ELISA.
 - This step can be done just prior to an ELISA or the day before (spun samples can be stored overnight at 4°C).

NOTE: If you are also going to prepare a tissue smear for the FAT, homogenize the tissue without dilution and make a smear from this material, then dilute the sample for the ELISA.

– Processing ovarian fluid.

- Dilute 1:2 (v/v); 1 volume ovarian fluid + 1 volume PBS-T20.
- Combine 1000 μL of ovarian fluid and 1000 μL of PBS-T20 in a microcentrifuge tube, store fluid or heat at 100°C for 15 min, remove debris by centrifugation (again, 8,000-10,000 x g for 10 minutes), and test supernatant.

– Processing plasma

- Adult fish: 1:5 (w/v); 1 volume whole blood + 4 volumes PBS-T20.
- Juvenile fish: 1:10 (v/v); 1 volume whole blood + 9 volumes PBS-T20.
- Collect blood in a syringe or capillary tube, then dispense the correct volume of blood into the appropriate volume of PBS-T20, remove the cellular fraction by low speed centrifugation, decant the supernatant, store fluid or heat at 100°C for 15 min, and test supernatant.

Appendix D: Description of controls

Control Group	Purpose	Coating Antibody	Sample Application	Step in the ELISA		
				Conjugate	Substrate-Chromogen	Stop Solution
Blank (BL)	Determine background absorbance levels in the absence of a test sample	Yes	PBS-T20 Only	Yes	Yes	Yes
Positive Controls (PC)¹	Internal control to ensure that predictable absorbance values are produced by certain levels of antigen	Yes	Yes	Yes	Yes	Yes
Negative Control (NC)²	Measure absorbance produced by sample from a negative control fish	Yes	Yes	Yes	Yes	Yes
Conjugate Control (CC)	Ensure that there is no nonspecific binding of the conjugate to well surfaces or to the coating buffer	Coating Buffer Only	PBS-T20 Only	Yes	Yes	Yes
Substrate-Chromogen Control (SC)	Test for nonenzymatic production of the color reaction	Coating Buffer Only	PBS-T20 Only	Diluent Only	Yes	Yes

¹Sample buffer containing a known level of *R. salmoninarum* (e.g. 1:100, 1:1,000, 1:2,000, and 1:5,000 (v/v) dilutions of control antigen).

²Appropriate tissues or body fluids with no, or undetectable levels of *R. salmoninarum* antigen. Store at -70°C; thaw morning of use; heat at 100°C for 15 min; and centrifuge 8,800 x g for 5 min.